



SPEED POST
Procurement Department
TENDER ENQUIRY

Ref: Q/NIB/17-18/86

Date: 18.07.2018

Due date of submission: 07.08.2018, Due Time: Up to 15:00 Hours
Due date of opening: 07.08.2018, Due Time: 15:30 hrs onwards.

Sub: - Tender Enquiry for Lab Equipment - Reg:-

Dear Sir,

I am to send you this Tender Enquiry with the request to submit the most competitive offer for supply of the following items as per details below.

S. No.	Description of Material	Quantity
1.	Gel Imager System with Accessories (Transilluminator)	01 No.
2.	Cellulose acetate Electrophoresis System (with Power supply) - Horizontal	01 No.

Please indicate price on following terms:

- Price:- FOR Destination NIB, NOIDA
- Taxes:- **GST percentage should be clearly mentioned. Otherwise, Bid will not be accepted.**
- Excise Duty: - NIB will provide necessary certificate as NIB is registered with DSIR for the purpose of availing Custom Duty exemption in terms of Government Notification No. 51/96 – Customs dated 23 July 1996 and Central Excise Duty exemption in terms of Government Notification No. 10/97 – Central Excise dated 1 March 1997 as amended from time to time. To avail the exemption, NIB shall be providing the above certificate along with the supply order. Subject to above, indicate applicable Duty.
- Payment: - 100% payment within 30 days from the date of satisfactory receipt and acceptance of material at NIB, NOIDA
- Delivery: - 4 weeks
- Earnest Money Deposit – The Bidder shall furnish a bid security of **Rs.1900/- (One thousand nine hundred only)** in the form of a Demand Draft drawn in favour of, "National Institute of Biologicals" payable at Noida/New Delhi or Bank Guarantee as per the enclosed format at Annexure – III from a Schedule Bank. The Bank Guarantee issuing branch should be situated in India and the Bank Guarantee

- should remain valid for a period of 45 days after the bid validity period. Demand Draft/BG is to be enclosed along with "Un – priced" bid only.
- vii) Any bid not accomplished by EMD will be rejected by the Purchaser as non responsive.
 - viii) Unsuccessful bidder's bid security will be discharged / returned as promptly as possible but not later than 30 days after the expiry of bid validity period prescribed by the purchaser.
 - ix) Transit Insurance – For delivery of goods at site, the insurance shall be obtained by the supplier for an amount equal to 110% of the value of goods from "warehouse to warehouse" (ultimate destination) on all risks" basis including strikes.
 - x) Delay delivery discount – In case of delayed delivery of all or any item (s) or goods the Purchaser shall be entitled to a discount or reduction on the contract price to a sum equivalent to 1% (one percent) of the value of goods/items delivered late for each week or part thereof subject to a maximum of 10% (ten percent) of undiscounted/unreduced contract price for the relative good/items, and/or, risk purchase will be affected at your risk and expenses.
 - xi) Warranty – One year from the date of supply against any manufacturing defects.

Terms and conditions for MSME

- i) In compliance of Public Procurement Policy for Micro and Small Enterprises (MSEs) Order 2012, purchase preference shall be given to Micro and Small Enterprises (MSEs) registered with District Industries Centers or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir Board or National Small Industries Corporation or Directorate of Handicrafts and Handloom or consortia of Micro and Small Enterprises formed by National Small Industries Corporation, if the price quoted by the MSEs is within 15% of L-1 price (L-1 price means Landed Lowest price) quoted by any other bidder, for the tendered items for which they are the L-1.
- ii) Purchase preference shall be applicable for the 20% of the tendered quantities only provided L-1 bidder is not MSE and the MSEs accept the L-1 price.
- iii) In case of an MSEs is the L-1 bidder for an item, the 100% tendered quantities shall be placed an order to MSEs firm.
- iv) In case of an MSE owned by SC/ST and price quoted is within 15% of the L-1 price quoted by the L – bidder, then 20% quantity shall be ordered on them at L-1 rate, subject to their acceptance.
- v) In case of more than one MSEs participates in tender and their prices are within 15% of L-1 bidder price who is not a MSEs and one of the MSEs is owned by SC or STY then 4% tendered quantities or say 20% of the 20% tendered quantities shall be procured from MSEs owned by SC or ST entrepreneurs exclusively subject to their acceptance of L-1 price in addition to sharing of equal portion of balance of 16% with other non-SC/ST owned MSEs subject to their acceptance of L-1.
- vi) In case of more than one MSEs owned by SC or ST then 4% tendered quantities or say 20% of the 20% tendered quantity shall be procured from these MSEs owned by SC or ST entrepreneurs exclusively subject to their acceptance.
- vii) In case of more than one MSEs owned by SC or ST then 4% tendered quantities or say 20% of the 20% tendered quantity shall be procured from these MSEs owned by SC or ST entrepreneurs subject to their acceptance.

Notes:

1. The offer should be submitted in two sets, One Set containing Technical details, Literature, Catalogue etc. and Tender commercial terms should be put in a sealed cover and marked as **'Technical Bid'**. Prices should be kept blank.
- 1.2 The Second set should be sent in a separate sealed cover and should contain Price Schedule only. **'Price – Bid'** should be clearly mentioned on the cover.
- 1.3 Tender No., Item description and Due Date of opening should be clearly mentioned on the Envelope.
- 1.4 Both the **'Technical – Bid'** & **'Price – Bid'** should then be put in a single sealed cover clearly indicating the Item, Tender No. & Due Date of opening.
- 1.5 Offer to be given by authorized dealer / distributor shall accompany with valid authorization Certificate issued by the principal. **If bidder has opted to quote more than one item and for each item their principal is different, then in such cases the bidder must submit authorization certificate for each quoted item from their principals. Please note that your offer will not be entertained without the valid authorization and the same will be summarily rejected.**
- 1.6 Copy of Sales Tax clearance certificate duly attested by a Gazette officer.
- 1.7 Customer reference list for the similar / materials for last two years.
- 1.8 Current valid certificate of registration including its annexure in case registered with SSI / NSIC etc.
- 1.9 Validity of the offer should be kept open for a period of not less than **three months**.

Please indicate item, our enquiry number and opening date on top of the envelope positively.

Yours faithfully



(W.Z. Quazi)
Procurement Officer

- Encl:
- 1) Annexure – II Price Schedule
 - 2) Annexure - III – Bid Security Form (EMD – BG)

Specification for Electrophoresis system with Power Supply - HORIZONTAL

Electrophoresis Unit:

- Principal Material should be Acrylic or High-transparent Polycarbonate, mold, hitting-proof, high temperature-proof, corrosion-proof.
- Inner Tank Dimension (L X W X H) should be minimum 22 x 15 x 9.5 cm.
- Number of combs should be minimum 2 Nos.
- Number of gel casting tray should be minimum 1 no.
- Connecting cord should be minimum 1 set (black and red).
- Number of platinum electrodes should be 1 anode and 1 cathode.
- Gel Dimensions should be 9 x 11 (cm) or 10 x 10 (cm) (gasketed).
- System should have comb thickness of 0.75, 1, 1.5 and 2 (mm)
- System should have combs of wells: 5, 8, 10, 12, 16, and 20 nos.
- System should have a buffer requirement minimum 250 (ml) & Max. Buffer volume 600 ml.
- All the accessories should be supply with main unit.
- IQ, OQ & PQ reports must be provided with the instrument.
- System should have 1 year warranty and CE-conform.

Power Supply Unit:

- Output: 4 output terminal
- Voltage: constant voltage 10 to 300V DC
- Current: constant current 4 to 400 mA
- Max. Power: 75 W
- Voltage Setting Resolution: min. 1V
- Current setting Resolution: min. 1mA
- Maximum Ripple: 1%
- Voltage Regulation: 1%
- Max. Operating Temp. 45°c ambient
- Display for voltage: 3 digit
- Display for current: 3 digit
- Timer: 1 minutes to 999 minutes
- Input Supply: 230 VAC
- Warranty: 1 year
- IQ, OQ & PQ reports must be provided with the instrument.

GEL IMAGER SYSTEM WITH ACCESSORIES
Specification for (UV Transilluminator)

- System should be ideal for RNA/ DNA samples visualization through UV Light.
- Working area should be minimum 15 x 15 cm or 20 cm X 20 cm independent for UV.
- 5 or 8 UV tubes of 312 nm as well as 365 nm wavelength fitted individually in tube holder.
- Intensity required minimum 5400/6400 $\mu\text{W}/\text{cm}^2$.
- Power Supply should be 220-240V, 50 Hz.
- UV tube average shelf life should be minimum 3000 Hrs.
- Availability of local service engineers and service facility would be preferred.
- List of installations of Transilluminator in India must be provided.
- IQ, OQ and PQ reports must be provided with the instrument.
- Warranty should be minimum 1 Year.