

राष्ट्रीय जैविक संस्थान

स्वास्थ्य एवं परिवार कल्याण मंत्रालय

NATIONAL INSTITUTE OF BIOLOGICALS

Ministry of Health & Family Welfare, Government of India

Q/NIB/17-18/116



A-32, Sector - 62 (Institutional Area)

NOIDA - 201309 (U.P.) - INDIA

Phones : 0120 - 2400022, 2400072

Fax : 0120 - 2403014, 2400074

e-mail : info@nib.gov.in Website : www.nib.gov.in

Dated: 02.05.2018

Sub: - Issue of Amendment to T.E # Q/NIB/17-18/116/5431 dated: 13.03.18 reg.

Sir/Madam,

I am to refer to our tender enquiry of even no. Q/NIB/17-18/116/5431 dated: 13.03.2018 on the above mentioned subject and to say that the due date for submission of tender is extended upto 11.05.2018.

Mentioned in Tender Enquiry.

S.No.	Description	Qty
1	Micro pH Electrode	01 No.

Please read as:

Description :-Micro pH Electrode			Qty
S.no.	Features	Specification.	
1.	Product Type	Micro pH Electrode	01 No
2.	Electrode Diameter	2.5 to 4mm	
3.	Electrode Length	130 to 180mm	
4.	Wire Length	1 meter	
5.	Material	Glass	
6.	Junction	Glass double junction	
7.	pH Range	1 to 14	
8.	Connector Types	Waterproof BNC Connector	

All other terms and condition mentioned in Tender Enquiry remains unaltered.

Yours faithfully,

(W. Z. Quazi)

Procurement Officer.